

AN001: PRESSURE COMPENSATION OF A CO₂ SENSOR


ABSTRACT

Non-dispersive infra-red (NDIR) sensors work by measuring the amount of IR light absorbed by the target gas in a fixed volume. Gas concentration is proportional to the amount of light absorbed as it passes through the gas.

The amount of light absorbed by the target gas, and hence measurement accuracy, is influenced by external environmental factors. The absorption of IR light is dependent on the number of gas molecules present in the fixed volume. Three factors influence the number of molecules in this fixed space, the gas concentration, gas temperature and gas pressure. To accurately measure gas concentration, it is therefore important to understand the effects of temperature and pressure.

The general behaviour of a gas, due to changes in temperature and pressure is based on well understood laws. However, the effects of temperature and pressure on the absorption of infra-red light by gas molecules and how they influence measurement accuracy are less well understood.

This application note describes the physical effects on the gas molecules due to changes in temperature and pressure. Changes in ambient temperature and pressure will induce CO₂ concentration measurement errors unless corrected.



It also explains how CO₂ sensor measurements are affected by temperature and pressure and how these effects can be compensated for to reduce their impact on sensor accuracy.

AN001: PRESSURE COMPENSATION OF A CO₂ SENSOR

TABLE OF CONTENTS

ABSTRACT.....	1
BASIC PRINCIPLES.....	3
ENVIRONMENTAL FACTORS.....	3
GAS VIBRATIONS.....	3
ABSORPTION	5
COMPENSATION OPTIONS.....	7
Temperature Compensation Techniques	7
Pressure Compensation Techniques.....	7
APPLYING PRESSURE COMPENSATION	8
ADDING EFFECTS OF SPECTRAL BROADENING	10
CONCLUSION.....	13
IMPORTANT NOTICE	14
ADDRESS	15
REVISION HISTORY	16

AN001: PRESSURE COMPENSATION OF A CO₂ SENSOR

BASIC PRINCIPLES

All NDIR sensors measure the concentration of CO₂ in a fixed volume, which in turn is based on the number of molecules present. The more molecules there are in the fixed volume, the more IR radiation is absorbed. Both temperature and pressure influence the number of molecules present in the fixed volume and therefore, measurement accuracy will be affected unless steps are taken to compensate for these effects.

ENVIRONMENTAL FACTORS

The two environmental factors that decide the number of molecules present in a defined volume for any gas are temperature and pressure. This is defined by the ideal gas law, which combines the work of Boyle, Charles & Gay-Lussac (from the 17th, 18th & 19th centuries).

The basic principles are simple to understand; as the gas temperature increases, the molecules vibrate more and in doing so, they take up more room, and the number present in a vented fixed volume is reduced. Similarly, if the pressure is increased more molecules are pressed together and give rise to a higher number in the same volume.

GAS VIBRATIONS

Before the additional effects of temperature and pressure are analysed, it is important to review why some CO₂ molecules absorb IR radiation and some do not.

CO₂ molecules have 4 degrees of vibrational freedom. The potential gas vibrations are defined as symmetrical, asymmetrical and two types of bending. The symmetric vibrations are due to the two oxygen atoms vibrating around the carbon atom while always being an equal distance from the carbon atom. Asymmetrical vibrations are due to the carbon atom moving back and forward between the 2 oxygen atoms. The two bending vibrations, so-called degenerate bends are less distinct, but effectively either vibrate up and down or backwards and forward.



AN001: PRESSURE COMPENSATION OF A CO₂ SENSOR



Figure 1: CO₂ Vibration Modes

The symmetric stretch does not create a dipole and consequently is not able to absorb any photons. It therefore does not absorb infra-red radiation. The other three vibration modes produce a dipole and absorb photons although at different frequencies. The asymmetric stretch absorbs radiation at 4.26 μ m, whereas the bending modes absorb at 15 μ m.

Most NDIR based sensors measure the absorption of photons at 4.26 μ m, and this is due to the asymmetric stretch vibration mode. The proportion of molecules vibrating in each mode remains the same as temperature and pressure are changed.

A typical GSS sensor will contain approximately 1×10^{16} CO₂ gas molecules at 25°C and 1013mbar (1 atmosphere pressure) at 400ppm concentration levels.

AN001: PRESSURE COMPENSATION OF A CO₂ SENSOR

ABSORPTION

Absorption at 4.26µm due to the symmetric vibration mode is negligible. This is because the average position of the gas molecule is the same as symmetric bending mode case, with no dipole and hence little absorption. However, as the atoms vibrate asymmetrically and form different dipole strengths depending on the energy levels of the electrons, the gas molecules begin to absorb infra-red radiation.

The different energy levels of the molecules cause absorption at a series of distinct wavelengths as shown below in Figure 2. The effects of changes to the absorption due to pressure and increased temperature scan be seen more clearly by focusing on one specific wavelength.




Figure 2: CO₂ Absorption Around 4.3µm

Figure 3 shows the impact of changes of pressure on absorption at 4.28µm. Figure 3 also shows how the absorption is affected by changing the gas pressure from 1.0 to 1.5 atmospheres, with all other parameters unchanged. Although the peak absorbance does not change significantly, the width of the peak increases as the pressure increases. This is caused by the increased collisions as the molecules vibrate, causing electron energy levels to vary and the gas absorbing photons over a slightly greater bandwidth.

AN001: PRESSURE COMPENSATION OF A CO₂ SENSOR




Figure 3: Absorption at 4.28µm vs. Pressure




Figure 4: Absorption at 4.28µm vs. Temperature

As the gas temperature is increased, the gas molecules vibrate more intensely and increase in energy. However, the absorption of infra-red radiation actually decreases, as shown in Figure 4 by the drop in the peak and narrowing of the absorption band. This is due to a reduction in the number of gas molecules in the fixed volume that are available to absorb the IR light.

AN001: PRESSURE COMPENSATION OF A CO₂ SENSOR

COMPENSATION OPTIONS

It has been shown that IR radiation is affected by temperature and pressure. Depending on CO₂ measurement accuracy requirements, it may be necessary to compensate for these effects.

Temperature Compensation Techniques

Changes in temperature effect sensor accuracy in several different ways. As well as those effects that can be described by simple gas laws, the LED and photo-diode, the electronics and the mechanical parts of the sensor will be all be affected by temperature.

There are several ways to mitigate the effects of temperature on sensor accuracy. Whilst not often practicable due to cost, power, and space concerns, controlling the temperature of the sensor such as with a thermoelectric cooler (TEC) can be an effective strategy.

Alternatively, the sensor accuracy variations due to changes in temperature can be characterised and corrected.

At GSS, each sensor has built-in temperature correction based on characterisation of its behaviour over temperature and gas concentration during factory calibration.

Pressure Compensation Techniques

The impact of pressure changes on sensor accuracy depends principally on the gas concentration range to be measured. For gas concentrations below 1%, the error induced on measurement accuracy for a change in pressure of 20mbar from the nominal 1013mbar reference level is less than 2%. In most so-called ambient applications, this is below the measurement accuracy of a typical sensor and well within the required measurement limits. Therefore, typically the effects due to pressure variations on measurement accuracy are ignored for concentrations below 1%.

Above 1% concentrations, and depending on the likely variation in pressure, there is often a need to compensate for the effects of pressure on sensor measurement accuracy. Left uncompensated, measurement accuracy will be impacted by over 30%, for a 200mbar change in pressure with 100% CO₂ gas.

AN001: PRESSURE COMPENSATION OF A CO₂ SENSOR**APPLYING PRESSURE COMPENSATION**

As discussed earlier, the shape of the absorption band changes with pressure mainly caused by increased molecule collisions. This effect is known as spectral broadening. As pressure increases, so do the number of collisions due to the higher density of molecules present, and as CO₂ has a higher density than air, there is also an effect driven by CO₂ concentration.

Ignoring the effect of spectral broadening, the measurements can be corrected for pressure based on basic gas laws. Boyle's Law states that pressure is inversely proportional to volume and is constant.

$$P1V1 = P2V2$$

If the pressure doubles, the volume must halve. This implies that for a fixed volume gas sensor, if the pressure doubles, there is twice the concentration of gas in the fixed volume of the sensor. The effects of pressure due to this basic gas law can be corrected using the following formula where P is the pressure in mbar.

$$\text{Corrected CO}_2 \text{ Value} = \frac{\text{CO}_2 \text{ Measurement Value from Sensor (C1)}}{1 - (1013 - P)/1013}$$

This can be re-written as follows where Y (Correction Factor) is -1/1013 or -0.000987.

$$\text{Corrected CO}_2 \text{ Value} = \frac{\text{CO}_2 \text{ Measurement Value from Sensor (C1)}}{1 + Y(1013 - P)}$$

AN001: PRESSURE COMPENSATION OF A CO₂ SENSOR

Figure 5 below shows how sensor measurements are affected using a constant Y correction factor. Without applying the correction, the sensor measurement would increase erroneously with increasing pressure. Including this correction factor maintains the accuracy of the measurement with changing pressure.




Figure 5: Effects of Pressure on Gas Concentration

AN001: PRESSURE COMPENSATION OF A CO₂ SENSOR

ADDING EFFECTS OF SPECTRAL BROADENING

As discussed earlier, the spectral broadening effect increases the impact of pressure as the increased concentration effect adds to the basic gas law. In the graph below, the compensation value Y combines both the spectral broadening and gas law effect to give an overall compensation value.




Figure 6: Correction Factor vs. Concentration

The correction factor has been derived from spectral data & confirmed experimentally. By inspection, at low gas concentrations and small variations in pressure from the nominal 1 atmosphere reference level, the impact on sensor accuracy is small.

The correction curve does not fit a simple algorithm but can be simplified for some applications depending on the concentration range to be measured. For concentrations above 30%, the coefficient can be simplified to a constant -0.0014. For other ranges, it is possible to select different coefficients depending on the range of the sensor and the nominal concentration being measured.

To improve the correction further, there are two further methods to consider. The correction factor is based on experimental data and can be used directly as part of a look up table. Alternatively, it is possible to describe the correction coefficient as a polynomial. Techniques for generating a polynomial fit are beyond the scope of this application note.

AN001: PRESSURE COMPENSATION OF A CO₂ SENSOR

Figure 7 below shows the best fit using a 6th order polynomial (orange line) compared to the theoretical correction coefficient (blue line).




Figure 7: Correction coefficient Y

The orange line is a 6th order polynomial fit described by the following equation.

$$Y = 2.811 \times 10^{-38} x C1^6 - 9.817 \times 10^{-32} x C1^5 + 1.304 \times 10^{-25} x C1^4 - 8.216 \times 10^{-20} x C1^3 + 2.311 \times 10^{-14} x C1^2 - 2.195 \times 10^{-9} x C1 - 1.471 \times 10^{-3}$$

Other methods can be used to generate a best fit curve including using a different number of terms in the polynomial, or other non-linear regression techniques. The typical trade-off to be made is the balance between accuracy and computational complexity.

AN001: PRESSURE COMPENSATION OF A CO₂ SENSOR

Figure 8 below shows the impact of using the polynomial fit correction curve versus uncorrected CO₂ measurements.




Figure 8: Comparison of Uncorrected and Corrected Measurement Errors vs. CO₂ Concentration

The corrected CO₂ measurement errors using the 6th order polynomial are typically <1% except at the extremes. This is compared to uncorrected errors of >43% for measurements at 800mbar and >20% for measurements at 1200mbar.

AN001: PRESSURE COMPENSATION OF A CO₂ SENSOR

CONCLUSION

Unless corrected, the accuracy of CO₂ measurements is compromised by the effects of ambient pressure changes, particularly at gas concentrations above 1%. This application note describes potential techniques for correcting for the effects of pressure on CO₂ measurement accuracy, using a constant correction coefficient, a look up table or using a polynomial fit. These techniques enable accurate pressure compensation to be applied to many different applications ranging from ambient CO₂ sensing to more demanding environments of up to 100% CO₂ concentration.

GSS can advise on the best approach to correcting errors due to changes in ambient pressure and provide detailed technical support data for each method described in this application note.

AN001: PRESSURE COMPENSATION OF A CO₂ SENSOR

IMPORTANT NOTICE

Gas Sensing Solutions Ltd. (GSS) products and services are sold subject to GSS's terms and conditions of sale, delivery and payment supplied at the time of order acknowledgement. GSS warrants performance of its products to the specifications in effect at the date of shipment. GSS reserves the right to make changes to its products and specifications or to discontinue any product or service without notice.

Customers should therefore obtain the latest version of relevant information from GSS to verify that the information is current. Testing and other quality control techniques are utilised to the extent GSS deems necessary to support its warranty. Specific testing of all parameters of each device is not necessarily performed unless required by law or regulation. In order to minimise risks associated with customer applications, the customer must use adequate design and operating safeguards to minimise inherent or procedural hazards. GSS is not liable for applications assistance or customer product design. The customer is solely responsible for its selection and use of GSS products. GSS is not liable for such selection or use nor for use of any circuitry other than circuitry entirely embodied in a GSS product.

GSS products are not intended for use in life support systems, appliances, nuclear systems, or systems where malfunction can reasonably be expected to result in personal injury, death or severe property or environmental damage. Any use of products by the customer for such purposes is at the customer's own risk.

GSS does not grant any licence (express or implied) under any patent right, copyright, mask work right or other intellectual property right of GSS covering or relating to any combination, machine, or process in which its products or services might be or are used. Any provision or publication of any third party's products or services does not constitute GSS's approval, licence, warranty, or endorsement thereof. Any third party trade-marks contained in this document belong to the respective third-party owner.

Reproduction of information from GSS datasheets is permissible only if reproduction is without alteration and is accompanied by all associated copyright, proprietary and other notices (including this notice) and conditions. GSS is not liable for any unauthorised alteration of such information or for any reliance placed thereon.

Any representations made, warranties given, and/or liabilities accepted by any person which differ from those contained in this datasheet or in GSS's standard terms and conditions of sale, delivery and payment are made, given and/or accepted at that person's own risk. GSS is not liable for any such representations, warranties, or liabilities or for any reliance placed thereon by any person.

AN001: PRESSURE COMPENSATION OF A CO₂ SENSOR

ADDRESS

Gas Sensing Solutions Ltd.
Grayshill Road
Cumbernauld
G68 9HQ
United Kingdom

AN001: PRESSURE COMPENSATION OF A CO₂ SENSOR

REVISION HISTORY

DATE	RELEASE	DESCRIPTION OF CHANGES	PAGES
03/05/2020	1.0	First revision	All